

See discussions, stats, and author profiles for this publication at: <https://www.researchgate.net/publication/339850875>

# BIOACTIVITY SAMPLES WITH $\text{SiO}_2 - \text{Y}_2\text{O}_3$ CERAMIC LAYERS PRODUCED BY SOL-GEL METHOD

Conference Paper · March 2020

CITATIONS

0

READS

43

2 authors:



[Anna Mazur-Nowacka](#)

Wrocław University of Science and Technology

24 PUBLICATIONS 98 CITATIONS

[SEE PROFILE](#)



[Badr Qasem](#)

Wroclaw University of Environmental and Life Sciences

14 PUBLICATIONS 85 CITATIONS

[SEE PROFILE](#)

# BIOACTIVITY SAMPLES WITH SiO<sub>2</sub>-Y<sub>2</sub>O<sub>3</sub> CERAMIC LAYERS PRODUCED BY SOL-GEL METHOD

ANNA MAZUR-NOWACKA<sup>1\*</sup>, BADR QASEM<sup>2</sup>

<sup>1</sup> DEPARTMENT OF ADVANCED MATERIALS TECHNOLOGIES,  
FACULTY OF CHEMISTRY, WROCLAW UNIVERSITY OF SCIENCE  
AND TECHNOLOGY, POLAND

<sup>2</sup> DEPARTMENT OF BIOORGANIC CHEMISTRY,  
FACULTY OF CHEMISTRY, WROCLAW UNIVERSITY OF SCIENCE  
AND TECHNOLOGY, POLAND

\*E-MAIL: ANNA.MAZUR-NOWACKA@PWR.EDU.PL

[ENGINEERING OF BIOMATERIALS 153 (2019) 74]

## Introduction

316L stainless steel is widely used in implantology, although biological complications may result from its insufficient mechanical properties and low corrosion resistance in the human body. In order to improve the corrosion resistance of 316L steel and its bioactivity, coatings are applied, e.g. ceramic layers that can improve the proliferation of living cells and barrier protection in the human body [1,2].

## Materials and Methods

14.8 mm diameter samples of 316L stainless steel were used as a substrate. The application of the layers was based on the sol-gel technique. The coating solutions were obtained from tetraethoxysilane (TEOS) and yttrium(III) nitrate. As the solvent, butanol was used. Acetic acid and nitric acid were used to accelerate the reaction. The coatings were deposited using the immersion method. They were deposited in various sequences on 316L steel.

The thickness of the oxide layers was measured using SEM/Ga-FIB scanning microscope. The surface topography and shape were analyzed by SEM and also the same for distribution and morphology of MG-63 cells in the surface after 24 and 96 hours. In additionally, the proliferation analysis was carried out by using cell counting every 24 hours. However, the cytotoxicity effect was analyzed by using MTT assay every 24 hours to determine the toxicity effect of coatings to MG-63 adherent cells.

## Results and Discussion

The photos showed correct morphology for MG-63 cells and their distribution was regular. The MTT test showed differences in cell proliferation on individual samples. Samples are not cytotoxic and have a good effect on cell proliferation. Electrochemical tests performed in SBF solution proved that the corrosion resistance of samples coated with SiO<sub>2</sub> and Y<sub>2</sub>O<sub>3</sub> increased compared to uncoated 316L steel.

## Conclusions

The produced sol-gel coatings of SiO<sub>2</sub>-Y<sub>2</sub>O<sub>3</sub> on 316L steel increase the barrier properties and bioactivity.

## References

- [1] Bih-Show L., Si-Bu W., Sheng-Bo H., Chaur-Jeng W., Jyh-Wei L. „Characterization of plasma polymerized organosilicon thin films deposited on 316L stainless steel”. *Thin Solid Films*, 660, 30, (2018) 637-645.
- [2] R.I.M. Asri, W.S.W. Harun, M. Samykano, N.A.C. Lah, S.A.C. Ghani, F. Tarlochan, M.R. Raza „Corrosion and surface modification on biocompatible metals: a review” *Material Science Engineering C*, 77, (2017) 1261-1274.